



TEST REPORT

Trans Pacific Textile Sdn. Bhd.
Test Report No:
7191243504-CHM20-02A-RC_CR1

Test Report

Company Name & Address	Honsin Apparel Sdn. Bhd. 531 Batu 2 ½, Jalan Kluang, 83000, Batu Pahat Johor Darul Takzim, Malaysia.
Product/Model(s)	ProXmask Filter
Standard/Specification(s)	Bacterial Filtration Efficiency ASTM F2101-19
Date of Test	2 September 2020
Test Report Number	7191243504-CHM20-02A-RC_CR1

This report is made on the basis of information and data from original report results.

Prepared by

Dr Zetty Azalea
R&D Chemist



TEST REPORT

Trans Pacific Textile Sdn. Bhd.
Test Report No:
7191243504-CHM20-02A-RC_CR1

1. Bacteria Filtration Efficiency (BFE) Test (ASTM F2101-19)

Test description:

ASTM F2101-19 standard test method for evaluating the bacteria filtration efficiency (BFE) of medical face mask materials, using a biological aerosol of *Staphylococcus aureus*.

Area contacting with the bacteria challenge: white side of the mask

Flow rate: 28.3 ± 0.3 L/min

Mean particle size of the challenge aerosol: $3 \mu\text{m} \pm 0.3 \mu\text{m}$

Test area: approximately 50 cm^2

Sample description:

ProXmask filter

Result:

Sample	BFE (%)	Average BFE (%)
1	>99.96	99.96
2	99.96	
3	>99.96	
4	>99.96	
5	>99.96	